PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Organization of:

ETALONET, S.A. DE C.V.

Rio Panuco 3508 Col. Villa los Pinos
Monterrey, Nuevo Leon, México. C.P. 64770

and hereby declares that the Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025:2017

Whereby, technical competence has been confirmed for the associated scope supplement, in the fields of:

Dimensional Inspection and Thermodynamic Testing

(As detailed in the supplement)

Accreditation claims for conformity assessment activities shall only be made from the addresses referenced within this certificate and shall apply solely to those activities identified in the related scope. This Accreditation is granted subject to the Accreditation Body rules governing the Accreditation referred to above, and the Organization hereby commits to observing and complying with those rules in their entirety.

For PJLA:

Tracy Szerszen
President

Perry Johnson Laboratory Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

Initial Accreditation Date:	*Issue Date:*	*Expiration Date:*
September 25, 2026	September 25, 2026	November 30, 2028

Accreditation No.:	*Certificate No.:*
73706	L26-856

The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: www.pjlabs.com

Certificate of Accreditation: Supplement

ETALONET, S.A. DE C.V.

Rio Panuco 3508 Col. Villa los Pinos
Monterrey, Nuevo Leon, México. C.P. 64770
Contact Name: Roberto Benitez. Phone: 818-398-2950

Accreditation is granted to the facility to perform the following conformity assessment activities:

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Thermodynamic	Refrigerator / Fridge Freezer, Oven, Incubator, Autoclave	Installation Qualification (IQ) Operating Qualification (OQ) Performance Qualification (PQ) Uniformity and Stability	Internal PPC-1-CAL-01 (based on IEC 60068-3-11 and NOM-059 SSA1 (section 9.5 to 9.6) and NOM-164 SSA1 (Section to 9)) Customer Requirements	Datalogger Omega	F1, F4	O
Thermodynamic	Constant Climate Chambers Environmental Warehouses Stabilities Chamber	Installation Qualification (IQ) Operating Qualification (OQ) Performance Qualification (PQ) Uniformity and Stability (Temperature and % Relativity Humidity)	Internal PPC-1-CAL-01(based on IEC 60068-3-11 and NOM-059 SSA1 (section 9.5 to 9.6) and NOM-164 SSA1 (Section to 9)) Customer Requirements	Datalogger Omega, Rotronic	F1, F4	O
Dimensional Inspection	Fixtures, Parts, Gauges and Tooling	GD&T	ASME Y14.5	PCMM	F1, F2	F, O

1. Location of activity:

Location Code	Location
F	Conformity assessment activity is performed at the CAB's fixed facility
O	Conformity assessment activity is performed onsite at the CAB's customer location

Certificate of Accreditation: Supplement

ETALONET, S.A. DE C.V.

Rio Panuco 3508 Col. Villa los Pinos
Monterrey, Nuevo Leon, México. C.P. 64770
Contact Name: Roberto Benitez. Phone: 818-398-2950

Accreditation is granted to the facility to perform the following conformity assessment activities:

2. Flex Code:

F0: When no flexibility is identified. There are no changes to items tested, characteristics identified or versions of methods except for updating to the most recent version of a standard method after verification.
F1: The laboratory has the capability to test a new item, material, matrix, or product similar in composition to item, material, matrix, or product identified on the scope
F2: The laboratory has the capability to introduce the newest revision of an accredited authoritative standard method (with no modifications) identified on the scope
F3: The laboratory has the capability to introduce a parameter/component/analyte to an accredited test method identified on the scope
F4: The laboratory has the capability to introduce a new revision of an accredited non-standard method using the same technology or technique identified on the scope
F5: The laboratory has the capability to introduce a validated method that is equivalent to an accredited method (using same technology or technique) identified on the scope for the same parameter, component, or analyte identified on the line item of the scope.